



# SmartLinx Neuron 3 Regulatory Notice

## Introduction

This Notice contains regulatory information for the SmartLinx Neuron 3 product and is therefore an extension of the *SmartLinx Neuron 3 Installation and Maintenance Guide*. It details the compliance statements that the product requires for its certification and approval. Capsule Technologies, Inc. is therefore committed to delivering products compliant with standards, laws and regulations.

## Certification Marks



Recognized component



Underwriters' Laboratories



Compliance with FCC Regulations  
Assembled from tested components  
Complete system not tested



IATA Lithium Battery Mark  
Hazard Warning Label (UN 3091)



IATA Lithium Battery Mark  
Hazard Warning Label (UN 3481)

## Regulatory compliance and approvals

### USA

#### Federal Communications Commission

SmartLinx Neuron complies with Part 15 of the FCC Rules. Operation is subject to two conditions:

- This device may not cause harmful interference.
- This device must accept any interference received, including interference that may cause undesired operation.

**Note:** This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.

- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help

**Note:** This device meets the requirements specified in Part 15E, Section 15.407 of the FCC Rules.

**Caution:** Capsule Technologies is not responsible for any radio or television interference caused by unauthorized changes or modifications to this equipment. Unauthorized changes or modifications could void the user's authority to operate the equipment.

**Caution:** Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate this equipment. This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.

#### Radiation Exposure Statement

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20 cm between the radiator and your body.

**Note:** The country code selection is for non-US model only and is not available on all US models. Per FCC regulation, all Wi-Fi products marketed in the USA must fixed to US-operation channels only.

### Canada

#### Innovation, Science and Economic Development Canada (formerly Industry Canada) Statement

CAN ICES-003 B / NMB-003 B

This device contains licence-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's licence-exempt RSS(s). Operation is subject to the following two conditions:

- (1) This device may not cause interference.
- (2) This device must accept any interference, including interference that may cause undesired operation of the device

#### Radiation Exposure Statement

This equipment complies with ISED radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20 cm between the radiator & your body.

**Caution:** The device operates in the band 5150-5350 MHz, for indoor use only.

For further information, contact your local Industry Canada office.

## Disposal

---

As you use the SmartLinx Neuron 3, you will accumulate solid wastes that require proper disposal or recycling. These include system components, batteries, and packaging materials.

### ***Recycling and the Environment***

Improper disposal of IT and medical equipment can have a negative impact on health and the environment. We recommend that you dispose of Capsule Technologies products, including all electronic devices, cables, batteries, and so on, at an appropriate facility to enable recovery and recycling. You can also recycle packaging and guides according to your local recycling regulations.

### ***Components***

At the end of its service life, the products described in this manual, as well as their accessories, must be disposed of in compliance with the guidelines regulating the disposal of such products. If you have questions concerning disposal of the product, contact Capsule Technologies or its representatives.

### ***Batteries***

The sealed, rechargeable batteries contain lead and can be recycled. Discharge (deplete) batteries prior to disposal. Place the battery in packaging which electrically isolates its contents. Do not puncture or place the battery in a trash compactor. Do not incinerate the battery or expose it to fire or high temperatures. Dispose in accordance with regional body-controlled guidelines and hospital protocol.

**Warning:** Improper disposal of batteries may create an explosion or contamination hazard. Always recycle batteries according to local regulations. Never dispose of batteries in refuse containers. Do not wrap them in metal or aluminum foil. Wrap them in newspaper before disposing of them. Do not burn them. Battery may explode if overheated.

### ***Packaging Material***

Retain original packaging materials for future use in shipping the system and its accessories. The recommendation includes corrugated shippers and inserts. Whenever possible, recycle the packaging of accessories.

## Standards Compliance

---

Field	Standard or regulation
<b>Safety</b>	IEC 60950-1 UL 2054 / IEC 62133 IEC 60601-1
<b>EMC</b>	IEC 60601-1-2 (Class B emission) IEC 61000-3-2 IEC 61000-3-3 CISPR 22 CISPR 32 CISPR 24 CISPR 35 IEEE ANSI C63.4
<b>Radio</b>	IEEE ANSI C63.10

### ***Essential Performance***

- The system maintains clinical data integrity up to delivery to a third-party system. This performance does not include: data not being delivered or the delay of data delivery.
- The system does not compromise the operation of connected medical devices (vital signs monitors).

**Legal Notice**

© 2019 Capsule Technologies, Inc. All rights reserved.

This product and related documentation are protected by copyright and distributed under licensing restricting their use, copying, distribution, and decompilation. No part of this product or related documentation may be reproduced in any form or by any means without prior written authorization of Capsule Technologies, Inc.

This technical data may be subject to U.S. and International export, re-export, or transfer (“export”) laws. Diversion contrary to U.S. and International laws is strictly prohibited.

SmartLinx Medical Device Information System and other associated trademarks or logos are trademarks or registered trademarks of Capsule Technologie SAS.

The hardware and software of Capsule Technologies, Inc. is provided subject to all third-party licenses and limitations in the “third party notices” electronic file included as part of the software or available upon request from Capsule Technologies, Inc.

All other trademarks, service marks, registered trademarks, and registered service marks are the property of their respective owners.

DHF18953

DCN 2019-002